EQUIPMENT REQUIREMENTS

As per Revised BDS Course Regulations, 2007

Department: Oral Medicine and Radiology

			ntity	Available	Adequate /
Name	Specification	50 Intake	100 Intake	Quantity	Inadequate
	(Space allocation for each dental chair 100 Sq.Ft. in all the Departments)				
De <mark>ntal Chairs</mark> an <mark>d</mark> Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventry Suction, X-Ray viewer, 3 way syringe, instrument tray, Dental Operator's Stool with height adjustment		12	12	Adequate
Panoramic X-Ray with Cephalometric	Preferably digital	1	1	1	Adequate
intra Oral X-ray Unit	70 KV, 8 mA, high frequency preferably Digital timer	2	3	3	Adequate
Pulp testers	Digital			2	Adequate
Automatic periapical X-rayDeveloper	Digital	1	1	1	Adequate
Automatic Panoramic with Cephalometric X-ray Developer		1	1	1	Adequate
X-ray viewer	For Panoramic and Cephalometric films	2	2	2	Adequate
Radiovisiograph y	Digital Intra X-ray system with one Sensor and Software	1	1	1	Adequate
General X-ray Unit		1	1	1	Adequate
Ortho Pantmograph	Preferably digital	1	1	1	Adequate
Automatic Developers / Dark Room		1	2	2	Adequate
Lead aprons		2	2	3	Adequate
Lead Gloves		1	1	1	Adequate
X-ray Hangers		6	6	6	Adequate
X-ray Viewers		2	2	2	Adequate
Diagnostic Kits	Mouth mirror, dental probe, college tweezers	20	40	40	Adequate
Lead Screen		1	1	1	Adequate
Biopsy Kit		1	1	1	Adequate
Autoclave	Having wet and dry cycle, which can achieve 135°Cwith minimum capacity of 20 liters	1	2	2	Adequate
Computers	Minimum Pentium – IV	1	1	1	Adequate
Intra Oral Camera	With High resolution	1	1	1	Adequate
Needle Burner With Syringe Cutter		2	2	Vardel	Adequate

Department: Oral Pathology and Oral Microbiology

Name	Specification	Quantity		Available	Adequate /
		50 Intake	100 Intake	Quantity	Inadequate
Dental Chairs and Units	With shadow less lamp, spittoon, 3 way syringe, instrument tray and suction	1	2	6	Adequate
Microscopes	Production and the second seco	20	40	89	Adequate
Microtome		1	1	1	Adequate
Wax bath		1	1	1	Adequate
Water bath		1	1	1	Adequate
		1	1	1	Adequate
Knife sharpener		1	1	1	Adequate
Hot plate Spencer Knife		1	1	1	Adequate

Department: Publ			ntity	Available	Adequate
Name	Specification	50 Intake	100 Intake	Quantity	Inadequate
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, micromotor control light cure 3 way syringe, X-ray viewer, instrument tray Dental Operators stool with height adjustment with shadowless lamp, spittoon, 3 way syringe, instrument tray and suction, micromotor, airotor, light cure	8	16	16	Adequate
Autoclaves	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	2	2	Adequate
Ultrasoniccleaner	Minimum capacity 13 litres with mesh bucket with digital timer	1	2	2	Adequate
Needle burner With syringe cutter		2	4	4	Adequate
MOBILECLINIC					Adequate
Mobile dentalvan	Mobile dental van with two dental chairs with all the attachments and adequate sitting space for 15 to 20 people			YES	Adequate
Dental chair with unit	Hydraulically operated with spittoon attachment, halogen light with 2 intensity, air ventury suction, air rotor, micromotor, 3 way-scaler and light cure, x-ray viewer, instrument tray, operating stool			YES	Adequate
Autoclaves	Having wet and dry cycle, which can achieve 135°Cwith minimum capacity of 20 liters	1	1	1	Adequate
Intraoral x-ray	Portable, 70 KV, 8 mA	1	1	1	Adequate
Glass bead sterilisers		1	1	1	Adequate
Compressor	1.25 HP	1	1	1	Adequate
Metal Cabinet	With wash basin	1	1	1	Adequate
Portable dental	Suitcase unit with airotor, micromotor, scaler and compressor 0.25 HP	2	2	2	Adequate
Stabiliser	4 KV	1	1	1	Adequate
Generator	4 KV	1	1	1	Adequate
Water tank	400 ltrs	1	1	1	Adequate
Oxygen cylinder		1	1	1	Adequate
Public address system		1	1	1	Adequate

TV and video	1	1	1	Adequate
cassette player	1	1	1	Adequate
Demonstration models				

Department:	Paediatric and Preventive Dentistry	Quai	ntity	Available	Adequate
me	Specification	50 Intake	100 Intake	Quantity	/ Inadequate
ental Chairs and nits	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power motorized evacuation system, Air rotor with miniature, Airrotor HPS, Micro motor with miniature contrangle Hand piece, 3 way syringe, ultrasonic sealer with 3 tips, Light cure unit LED based heat free, X-ray viewer, instrument tray Dental Operator'sstool with height adjustment (Pedo chair preferred)	10	20	20	Adeq uate
utoclaves	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	2	2	Adequate
Iltrasoniccleaner	Minimum capacity 13 liters with mesh bucket withdigital timer	1	2	2	Adequate
Veedle Burner With syringe		2	3	3	Adequate
cutter		1	1	1	Adequate
Amalgamator		1	1	1	Adequate
Pulp TesterDigital Rubber dam kit		3	5	5	Adequate
for pedo		1	1	1	Adequate
Apex locator	With torque control HPs	1	1	3	Adequate
Endo motor Injectable gutta percha wit condensation	a h	1	1	1	Adequate Adequate
Radiovisiography	Digital intra X-ray system with pedo sensor and software	1		1	Adequate
Intra Oral Camer	a With high resolution	1	1		Adequate
Scale instrument		5	10	10	Adequate
Restorative instruments		5	10	10	
Extractionforcep	ne l	5	10	10	Adequate
Intra-oral X-ray		1	1	1	Adequate
Automatic Developer		1	1	1	Adequate Adequate
Computer	Minimum Pentium - IV	1	1	1	Adequate
Computer					
	PEDO LAB		2	2	Adequate
Plaster dispense	one each for plaster and stone plaster	2			Adequate
Model Trimmer	With diamond disc	1			Adequate
Model Trimmer	Double disc one diamond and one carborandam disc	1			Adequat
Welder w soldering attachments	rith				
	The second secon	2	2 2		Adequat
Vibrator	or Heavy duty with Hps		2 3		Adequat
Lab micro moto Dental Lathe	lobout 10 lobout		1 1		Adequat
Dental Latne			1 1	1	Adequat

Model Trimmer	N. C. S.	1	1	1	Adequate
Steam cleaner		1	1	1	Adequate
Pressure moulding machine		1	1	1	Adequate
Carborandum disc		1	1	1	Adequate
Diamond disc		1	1	1	Adequate

Department: Orthodontics

	Specification	Qua	ntity	Available	Adequate Inadequate
Name		50 Intake	100 Intake	Quantity	
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, micromotor control light cure 3 way syringe, X-ray viewer, instrument tray Dental Operators stool with height adjustment	9	18	18	Adequate
Unit mount scaler		3	5	5	Adequate
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	2	3	Adequate
Ultrasoniccleaner	Minimum capacity 13 litres with mesh bucket with digital timer	1	2	2	Adequate
ORTHO LAB					Adequate
Plaster dispenser	One each for plaster and stone plaster	2	2	2	Adequate
Vibrator		2	2	2	Adequate
Model Trimmer		1	2	2	Adequate
Micromotor	Heavy duey	2	4	5	Adequate
Lathe		1	2	2	Adequate
X-ray viewers		2	2	3	Adequate
	If available in radiology its is adequate	1	1	Yes	Adequate
Welders		2	4	5	Adequate
Blue Torch		1	1	1	Adequate
Base Formers		2	4	4	Adequate
Typodont		2	4	4	Adequate
Set of pliers		5	10	11	Adequate
Welder with soldering attachments		1	1	1	Adequate
Hydro solder		1	1	1	Adequate
Typhodont articulator	With metal teeth wax rim of Class I, II, III	3	4	4	Adequate
Pressure moulding machine		1	1	1	Adequate

Madelal

Department: Periodontology

Department: Periodontology		Quantity		Details	Adequate /
Na <mark>me</mark>	Specification	50	100 Intake	Website	Inadequate
and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, with shadowless lamp, spittoon, 3 way syringe, Ultrasonic sealer with 3 tips, X-ray viewer, instrument tray Dental Operators stool with height adjustment		34		Adequate
Scaling	Operators stool with height asjactment	5	10	10	Adequate
instrument sets Surgical		3	6	6	Adequate
instrument sets	the suple which can achieve	2	3	3	Adequate
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	-			
Ultrasonic	Minimum capacity 13 litres with mesh bucket	1	2	2	Adequate
scaler Electrosurgical		1	1	3	Adequate
Needle burner withsyringe cutter		4	6	6	Adequate
LASER	Soft tissue laser	1	1		Adequate
Surgical moto With physio dispenser		1	1	2	Auequate

Department: Conservative Dentistry and Endodontics

Department: Conservative Dentistry and Endoughties		Quantity		Details	Adequate /
Name	Specification	50 Intake	100 Intake	Website	Inadequate
and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, with shadowless lamp, spittoon, 3 way syringe, Ultrasonic sealer with 3 tips, X-ray viewer, instrument tray Dental Operators stool with height adjustment		34	35	Adequate
1 1/1-		4	6	6	Adequate
Rubber dam kits		5	10	10	Adequate
Restorative instruments kits		5	10	10	Adequate
R.C.T. instrument					Adequate
kits Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	2	3	3	
I Illiana a min	Minimum capacity 13 litres with mesh bucket	1	2	2	Adequate
Ultrasonic cleaner		3	4	4	Adequate
NeedleBurner withSyringe Cutter	Madelcer				

		2	3	3	Adequate
Amalgamator	With auto proportion, auto dispenser	4	6	6	Adequate
Rubber dam kit	s	2	4	4	Adequate
Pulp Tester Digital		1	2	2	Adequate
Apex locator		4	6	6	Adequate
Glassbead					

					Adequate
terilizers		2	2	2	Adequate
Plaster			2	2	Adequate
lispenser		1	2	1	Adequate
/ibrator		1	1		Adequate
Ceramic Unit		1	1	1	Adequate
Casting machine	V. C cofoty	1	1	1	Adequate
Intra-oral X-ray Proper ra Unit	diation salety	1	1	1	Adequate
Automatic				1	Adequate
Developer	Computer	1	1	1	Aucquate
Radiovisiograph RVG with	1 Compute.			1	Adequate
y NASH tora	ue control Hps	1	1		Adequate
	de control ripo	1	1	1	
Bleaching Unit		1	2	2	Adequate
Magnification					
loops		1	2	2	Adequate
Injectable gutta					
		ating Light 30	60	60	Adequate
PHANTOM LAB Phantom Phantom movement Micro mb	Table fitted with Halogen Opera Head body type neck joint fint, TMJ movement, Modular with cotor with contra angle Hps, 3 was ivorine teeth, preferably soft ging as stool (not to use extracted or care	Air rotor, y syringe, val, dental			

DEAN
W.T. Dental Coll

Ch

mical Laboratory		Quantity		Available	Adequate
Name	pecification	50 Intake	100 Intake	Quantity	/ Inadequat
	l to an alastar	2	2	2	Adequate
laster	One each for plaster and stone plaster				1 1 mate
ispenser	1 11-4	1	1	1	Adequate
Model Trimmer	Carborandum disc	1	1	1	Adequate
	Diamond disc	1	2	2	Adequate
athe	Heavy Duty	2	3	3	Adequate
ab Micromotor	With heavy duty handpiece	1	1	1	Adequate
Ultrasonic	Minimum capacity 5 litres				1
cleaner		1	1	1	Adequate
Spindle Grinder		1	2	2	Adequate
Vibrator		1	1	1	Adequate
Burnout furnace		1	1	1	Adequate
Porcelain furnace		1	1	1	Adequate
Sandblasting Machine		1	1	1	Adequate
Lab Airotor		1	1	1	Adequate
Pindex System		1	1	1	Adequate
Circular saw		1	1	1	Adequate
Vacuum mixer		1	1	1	Adequate
Pnematic chise	Motor cast / induction casting preferred	1	1	1	Adequate

Mandelas

Department: Oral & Maxillofacial Surgery

11	Evodontia	

A) Exodontia		Qua	ntity	Available	Adequate /	
Name	Specification	50	100 Intake	Quantity	Inadequate	
and Units I	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, Air Ventury suction, X-ray viewer, 3 way syringe, instrument tray Dental Operators stool and height adjustment and suction,		30	30	Adequate	
Autoclave	Micromotor / Airotor Having wet and dry cycle, which can achieve 135°C	2	2	2	Adequate	
Ultrasonic	with minimum capacity of 20 liters Minimum capacity 13 litres with mesh bucket	1	2	2	Adequate	
cleaner		4	6	6	Adequate	
Needle Burner With Syringe					Adequate	
Cutter Extraction	Complete set	10	20	20		
forceps sets		5	10	10	Adequate	
Dental elevators Minor Oral	Complete set	3	6	10	Adequate	
Surgery kits Emergency		1	1	1	Adequate	
drugs tray		1	1	1	Adequate	
Oxygen cylinder with mask		2	2	2	Adequate	
X-ray viewers		1	1	1	Adequate	
Computer	Minimum Pentium – IV	1	1 1			

B) Minor Surgery

B) Minor Su	irgery	Qua	ntity	Available	
Name	Specification	50 Intake	100 Intake	Quantity	/ Inadequat
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evaculation system, Air ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contra-angle Handpiece, 3 way syringe instrument tray Dental Operators stool and height adjustment and suction.	i	5	5	Adequate

D:\LIC_2023-24\LIC_Forms\LIC_Form_Dental_25012023.docx

Department: Prosthodontics and Crown & Bridge

Name	Specification		intity	Available Quantity	Adequate /Inadequat e
		50 Intake	100 Intake		
Dental Chairs and Units	Electrically operated, Spittoon attachment, Halogen Light with 2 intensity, high power evacuation system, Air Ventury suction, X-ray viewer, Airrotor, Micromotor with straight and contrangle Handpiece, 3 way syringe, Ultrasonic sealer with 3 tips, Light cure unit, instrument tray and suction, Dental Operators stool with height adjustment		34	34	Adequate
Semi-adjustable articulator	With face bow	2	2	4	Adequate
Extra oral / intra oral tracer		2	2	4	Adequate
Dewaxing unit		1	2	2	Adequate

Curing unit		1	2	2	Adequate
Dental casting		1	1	1	Adequate
machine					
Wax burnout		1	1	1	Adequate
furnace					
Preheating		1	1	1	Adequate
furnace					
Surveying unit		1	2	2	Adequate
Heavy duty hand	Lab micromotors	3	4	4	Adequate
piece					
Autoclave	Having wet and dry cycle, which can achieve 135°C with minimum capacity of 20 liters	1	2	2	Adequate
Needle Burner		1	2	2	Adequate
With Syringe					
Cutter					
	One each for plaster and stone plaster	2	2	2	Adequate
dispenser					
ModelTrimmer		1	1	1	Adequate
with					
Carborandum					
disc					
Model Trimmer		1	2	2	Adequate
withdiamond					
disc					
Acryliser		2	3	3	Adequate
Lathe		1	2	2	Adequate
Flask press		4	4	4	Adequate
Deflasking unit		4	4	4	Adequate
Dewaxing unit		2	3	3	Adequate
Hydraulic Press		2	3	3	Adequate
Mechanical		1	2	2	Adequate
Press					
Vacuum mixing		1	1	1	Adequate
machine					
	With heavy duty handpiece	3	4	4	Adequate
Curing pressure		1	1	2	Adequate
pot					
Porcelain		1	2	2	Adequate
furnace					
Vibrator		1	2	2	Adequate
Sandblasting unit		2	2	2	Adequate
Ultrasonic		1	2	2	Adequate
cleaner					
Model Trimmer		2	4	4	Adequate
Hot		1	2	2	Adequate
wate					
r					
sterilizer		19			
Geyser	Compound bath	1	2	2	Adequate
H.P. grinder with suction		2	3	3	Adequate
Heavy duty lathe		2	2	2	Adequate
Phantom heads		50	50	50	Adequate
Pre-clinical	Gas connection & Bunsen burner	50	100	100	Adequate
working tables					

D:\LIC_2023-24\LIC_Forms\LIC_Form_Dental_25012023.docx

DEAN Y.M.T. Dental College & Hospital Kharghar, Navi Mumba' - 410 210

MNadekar

Ceramic and Cast Partial Laboratory

N.		Quantity		Available	Adequate
Name	Specification	50 Intake	100 Intake	Quantity	/ Inadequa
Plaster dispenser	One each for plaster and stone plaster	2	2	2	Adequate
Duplicator		1	1	1	Adequate
Pindex System		1	1	1	Adequate
Circular saw		1	1	1	Adequate
Burnout furnace		1	1	1	Adequate
Sandblasting Machine	With two containers	1	1	1	Adequate
Electro polisher		1	1	1	Adequate
Model Trimmer with Carborandum disc		1	1	1	Adequate
Model Trimmer with diamond disc		1	1	1	Adequate
Induction casting machine		1	1	1	Adequate
Programmable porcelain furnace with vacuum pump with instrument kit and material kit		1	1	1	Adequate
Spot welder with soldering attachmentof cable		1	1	1	Adequate
Vacuum mixing machine		1	1	1	Adequate
Steam Cleaner		1	1	1	Adequate
Spindle Grinder 24000 rpm with vacuum suction		1	1	1	Adequate
Wax heater		1	1	1	Adequate
Wax carver		1	1	1	Adequate
Curing pressure pot		1	1	1	Adequate
Milling machine		1	1	1	Adequate
Heavy duty lathe with suction		1	1	1	Adequate
Preheating furnace		1	1	1	Adequate
Palatal trimmer		1	1	1	Adequate
Ultrasonic cleaner	5 liters capacity	1	1	1	Adequate
Composite curing unit	o mers capacity	1	1	1	Adequate
Micro surveyor		1	1	1	Adequate
Pre-clinical Prosthetic Laboratory	Work table preferably complete stainless steel	30	60	60	Adequate
	fitted with light, Bunsen burner, air blower, working stool. Adequate number of lab micro motor with attached hand piece.		20	20	
Plaster room for pre-clinical work					Adequate
Plaster dispenser	One each for plaster and stone	2	2	2	Adequate
Vibrator		2	2	2	Adequate
Lathe	COLUMN SECTION OF	2	2	2	Adequate
Model Trimmer		1	1	1	Adequate
Carborandum Disc		1	1	1	Adequate
Diamond disc		1	1	1, 1, 1	Adequate

D:\LIC_2023-24\LIC_Forms\LIC_Form_Dental_25012023.docx

Model trimmer with diamond disc Model trimmer with double disc (one		1	1	1	Adequa
Carborandum and one diamond disc)		1	1	1	Adequa
Casting machine, motor cast with the safety door closure, gas blow torch with regulator		1	1	1	Adequa
Dewaxing furnace					
Induction casting machine with vacuum		1	1	1	Adequa
pump, capable of casting titanium chrome cobalt precision metal		1	1	1	Adequa
Spot welder with soldering, attachment cable	of	1	1	1	Adequat
Steam cleaner		1	1	1	Adaguat
Vacuum mixing machine		1	1	1	Adequat Adequat
Spindle grinder 24,000 ROM with vacuum suction		1	1	1	Adequat
Wax heater		2	3	4	Adequat
Wax carvers (Full PKT Set)		2	3	5	Adequat
Milling machine		1	1	2	Adequate
Stereo microscope		1	1	2	Adequate
Magnifying work lamp		1	1	3	Adequate
Heavy duty lathe with suction		1	1	1	Adequate
Preheating furnace Dry model trimmer		1	1	1	Adequate
Die cutting machine		1	1	2	Adequate
Ultrasonic cleaner		1	2	2	Adequate
Composite curing unit		1	1	1	Adequate
		1	1	1	Adequate
Ceramic Lab Equipm Fully programmable porcelain furnace	nent				
with vacuum pump		1	1	1	Adequate
Ceramic kit (instruments)		3	3	6	Adamid
Ceramic materials (kit)		1	1	2	Adequate
Ceramic polishing kit		2	2	2	Adequate Adequate
Implant Equipmen	it		-		Auequate
Electrical dental chair and unit		1	1	1	Adequate
Physio dispenser		1	1	1	Adequate
mplant kit	Minimum 2 systems	2	2	2	Adequate
mplants		10	10	89	Adequate
Prosthetic components		10	10	20	Adequate
Unit mount light cure		1	2	2	Adequate
K-ray viewer Needle destroyer		1	2	2	Adequate
		1	2	2	Adequate
Ultrasonic cleaner capacity 3.5 Its		1	1	1	Adequate
autoclave programmable for all ecommended cycles -ray machine with RVG		1	2	2	Adequate
Refrigerator		1	1	1	Adequate
urgical kit/prosthetic kit		1	1	1	Adequate
ducating models		2	2	2	Adequate
mplant removing instruments		1	1	5	Adequate
C_2023-24\LIC_Forms\LIC_Form_Dental_25012023.		1	1	1	Adequate

EQUIPMENT REQUIREMENTS

As per MDS Course Regulation, 2017

Department: Prosthodontics and Crown & Bridge

Name	Specification	Quantity		Available Quantity	
Electrical Dental Chairs and Units		unit p studen two ch	per PG	19	Adequate
Articulators - somi adjustable (- 1: - 1)		1 Unit	2 Units		
Articulators - semi adjustable/ adjustable with face bow		6	12	16+1	Adequate
Micromotor - (Lab Type can also be attached (fixed) to wall		2	4	8	Adequate
Ultrasonic scaler		2	2	6	Adequate
Light cures		2	2		Adequate
Hot air oven		1	1		Adequate
Autoclave		2	2		Adequate
Surveyor		2	2		Adequate
Refrigerator		1	1		Adequate
X-ray viewer		1	2		Adequate
Pneumatic, Crown bridge remover		2	3		Adequate
Needle destroyer		1	2		Adequate
Intra oral camera		1	1		
Digital SLR camera		1	1		Adequate
Computer with internet connection with attached printer and scanner	40716 4563	1	1		Adequate Adequate
LCD projector		1	1	1	Adequate
Clinical Lab for Prosthe	tics	*	-	1	ruequate
Plaster dispenser		2	2	2	Adaquata
Model trimmer with carborandum Disc		1	2		Adequate Adequate
Model trimmer with diamond disc		1	2		Adequate
High speed lathe		2	3		dequate
/ibrator		2	4		dequate
Acrylizer		1	2		dequate
Dewaxing unit		1	2		dequate
Hydraulic press		1	1		dequate
Mechanical press		1	1		dequate
acuum mixing machine		1	1		dequate
Micro motor lab type		2	3		dequate
uring pressure pot		1	1		dequate
ressure molding machine		1	1		dequate
Chrome - Cobalt Lab Equip	ment	1	-	- ^	uequate
uplicator		1	1	1 A	dequate
index system		1	1		dequate
urn-out furnace		1	1		dequate
Velder		1	1		dequate
andblaster	Micro and macro	1	1		dequate
lectro – polisher		1	1		dequate
lodel trimmer with carborandum disc C_2023-24\LIC_Forms\LIC_Form_Dental_25012023		Anlac			dequate

D:\LIC_2023-24\LIC_Forms\LIC_Form_Dental 25012023.docx

Department: Periodontology

Name	Specification	ne Specification Qua	Quantity		Available Quantity	Adequate Inadequate
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument tray and motorized suction, micromotor attachment with contra angle hand piece, airoter attachment, ultrasonic scaler (Piezo) with detachable autoclavable hand piece	One chair and unit per post graduate student and Two chairs with unit for the faculty		19	Adequate	
		1 Unit	2 Units			
Auto clave (fully automatic) front loading		1	2	3	Adequate	
Steel bin		4	6	21	Adequate	
Airoter hand pieces		2	2	7	Adequate	
JV chamber		1	1	1	Adequate	
Formalin chamber		1	1	1	Adequate	
W.H.O probe		2	2	12	Adequate	
Nabers probe		2	2	14	Adequate	
Williams probe		2	2	14	Adequate	
UNC-15 probe		4	4	4	Adequate	
Gold Man fox probe		1	1	3	Adequate	
Pressure sensitive probe		1	1	1	Adequate	
Marquis color coded probe		1	1	2	Adequate	
Supra gingival scalers	set	2	2	10	Adequate	
Sub gingival scaler	set	2	2	10	Adequate	
Arkansas sharpening stone	300	1	1	1	Adequate	
Surgical Instrume	nts	1				
Routine surgical instrument kit (Benquis periosteal elevator, periotome)	set	2	3	6	Adequate	
Surgery trolleys		6	6	14	Adequate	
X ray viewer		1	2	5	Adequate	
Surgical cassette with sterilisation pouches		4	6	10	Adequate	
Electro surgery unit		1	1	3	Adequate	
Special Surgical Instrui	nents					
Kirkland's knife	set	1	1	2	Adequate	
Orban's knife	set	1	1	4	Adequate	
Paquette blade handle		1	1	2	Adequate	
Krane kaplan pocket marker	set	1	1	4	Adequate	
Mc Calls universal curettes	set	1	1	1	Adequate	
Gracey's curettes (No.1-18)	set	2	2	7	Adequate	
Mini five curettes	set	1	1	1	Adequate	
Cumine scalar		1	1	2	Adequate	
Mallet		1	1	3	Adequate	
Chisel		1	1	3	Adequate	
Oschenbein chisel	straight, curved	1	1	2	Adequate	
Schluger bone file		1	1	8	Adequate	
Bone fixation screw kit		1	1	1	Adequate	
Bone scrapper		1	1	2	Adequate	
Bone trephines for harvesting autografts	1 set	1	1	1	Adequate	
Bone regenerative materials	Bone graft and GTR membranes	5.1	bo 15	7	Adequate	

Bone regenerative materials	Bone graft and GTR membranes	5	5	7	Adequate
Local drug delivery systems	At least two different agents	1 each	1	2	Adequate
Root conditioning agent	At least two different agents	2	2	2	Adequate
Micro needle holder		1	1	1	Adequate
Micro scissors		1	1	1	Adequate
Magnifying loop (2.5 – 3.5)		1	2	2	Adequate
Operating microscope	optional	1	1	0	Adequate
3rd generation digital probe	optional	1	1	0	Adequate
Bone expander and bone crester	optional	1	1	0	Adequate
Distraction osteogenesis kit	optional	1	1	0	Adequate
Bone mill	optional	1	1	0	Adequate
Bone graft / membrane placement spoon		1	1	1	Adequate
Bone condenser		1	1	1	Adequate
Peizo-surgery unit	optional	1	1		Adequate
Centrifuge for PRP/PRF preparation	optional	1	1		Adequate
Soft tissue laser (8 watt)		1	1	1	Adequate
Osteotome	set optional	1	1	2SETS	Adequate
Miscellaneous Instrume		-	-	20010	racquate
Composite gun with material kit		1	1	1	Adequate
Splinting kit with material		2	3	3	Adequate
Composite finishing kit		1	1	1	Adequate
Glass Ionomer cement		1	1	1	Adequate
Digital camera		1	1	1	Adequate
Intra Oral camera		1	1	1	Adequate
Ultrasonic cleaner		1	1	2	Adequate
Emergency kit		1	1	1	Adequate
Refrigerator		1	1	1	Adequate
X-ray viewer		2	2	5	Adequate
LCD projector				1	Adequate
Computer with internet connection with attached printer and scanner		1	1	3 COMPUTER S 1 PRINTER	Adequate
Implant Equipment				6	Adequate
Electrical dental chair and unit		1	1	3	Adequate
Physio dispenser		1	1	1	Adequate
Implant kit	At least two different systems	2	2	2	Adequate
Implants		10	10	18	Adequate
Implant maintenance kit (plastic instruments)		1 set	1 set	1	Adequate
Implant guide		1	1	1	Adequate
X-ray viewer		1	2	5	Adequate
Needle destroyer		1	2	6	Adequate
Ultrasonic cleaner capacity 3.5 lts		1	1	1	Adequate
Autoclave programmable for all recommended cycles		1	1	3	Adequate
RVG with x-ray machine		1	1	1	Adequate
Refrigerator		1	1		Adequate
Surgical kit		2	2		Adequate
		1	1		Adequate
Sinus lift kit					
Sinus lift kit Educating models		1	1	1	Adequate

Name	Specification	Quantity		Available Quantity	Adequate / Inadequa e
Dental Chairs and Units	with shadowless lamp, spittoon, 3 way syringe, instrument tray and high otorized T			19	Adequate
Autoclave			2 Units		
	Front loading	2	3	3	Adequate
Fumigators		1	1	2	Adequate
Oscillating saw	With all hand pieces	1	1	1	Adequate
Surgical Instruments					
General surgery kit including tracheotomy kit		2	2	2	Adequate
Minor oral surgery kit		5	10	10	Adequate
Osteotomy kit		1	1		Adequate
Cleft surgery kit		1	1		Adequate
Bone grafting kit		1	1		Adequate
Emergency kit		1	1		Adequate
Trauma set including bone		2	2		Adequate
Plating kit		1	1		Adequate
Implantology kit	Minimum 2	2	2		Adequate
With implants	systems	10	10		Adequate
Distraction osteogenesis kit		1	1		Adequate
Peizo surgical unit		1	1		Adequate
Magnifying loops		1	1		Adequate
Operating microscope and Microsurgery kit	desirable	1	1		Adequate
Dermatomes		1	1	2	Adequate
Needle destroyer		2	3		Adequate
Ultrasonic Cleaner capacity 3.5 Its		1	1		Adequate
ormalin chamber		1	1	2	Adequate
Pulse oxymeter		1	1		Adequate
Ventilator		1	1		Adequate
Major operation theatre with all facilities		1	1		Adequate
Recovery and Intensive Care Unit with all necessary life support equipments		2 beds	2 beds		Adequate
ibrooptic light		1	4	1	\daa
npatient beds		20	1		Adequate
Fiber optic laryngoscope		1	20		Adequate
Computer with internet			1		Adequate
connection with attached printer and scanner		1	1	2	Adequate
.CD projector		1	1	1 4	dequate
Refrigerator		1	1		dequate
		1	1	las	uequate

D:\LIC_2023-24\LIC_Forms\LIC_Form_Dental_25012023.docx

Department: Conservative Dentistry and Endodontics

lame	Specification		Quantity		Adequate / Inadequat e
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument tray and motorized suction, micromotor, airotor attachment with hand pieces	unit per gradua student two cha	post- te and airs	20	Adequate
		1 Unit	2 Units		
ENDOSONIC HANDPIECES – Micro endosonic Tips, retro treatment		2	3	3	Adequate
Mechanised rotary instruments including hand pieces (speed and torque control) and hand instruments various systems		3	6	6	Adequate
Rubber dam kit		1 per chair	1 per chair	20	Adequate
Autoclaves for bulk instrument sterilization vacuum (Front loading)		2	3	3	Adequate
Autoclaves for hand piece sterilization		1	1	6	Adequate
Apex locators one for every two chairs		2	4	6	Adequate
Pulp tester		2	4	4	Adequate
Equipments for injectable thermoplasticized gutta percha		1	2	2	Adequate
Operating microscopes 3 step or 5 step magnification		1	1	1	Adequate
Surgical endo kits (Microsurgery)		2	2	2	Adequate
Set of hand instruments (specifications required)	Alexander of	1	2	2	Adequate
Sterilizer trays for autoclave		4	4	6	Adequate
Ultrasonic cleaner capacity 3.5 Its		1	1	1	Adequate
Variable Intensity polymerization equipments - VLC units	Desirable	1	1	1	Adequate
Conventional VLC units one for every two chairs		2	4	4	Adequate
Needle destroyer		2	2	2	Adequate
Magnifying loupes one for students and one for faculty		1	2	2	Adequate
LCD projector		1	1	1	Adequate
Composite kits with different shades and polishing kits		2	4	4	Adequate
Ceramic finishing kits, metal finishing kit	s In ceramic labs	2	3	3	Adequate
Amalgam finishing kits		2	3	3	Adequate
RVG with x-ray machine developing kit	000	1	1	1	Adequate
Chair side micro abrasion	Mandella	1	1	1	Adequate

Bleaching unit	1	1	1	Adequate
Instrument retrieval kits	1	1	1	Adequate
Computer with internet connection with attached printer and scanner	1	1	1	Adequate
Refrigerator	1	1	1	Adequate
Equipments for casting procedures				Adequate
Equipments for ceramics including induction casting machines/ burnout preheat furnaces/ wax elimination furnaces	1	1	1	Adequate
Lab micro motor/ metal grinders / sand blasters/ polishing lathes/ duplicator equipment/ vacuum investment equipments	1	1	1	Adequate
Laser (preferably hard tissue)	1	1	1	Adequate
Face bow with semi adjustable articulator	1	2	2	Adequate

Dtmont	Paediatric and	Preventive	Dentistry
Donartment	Paediatric and	Preventive	Dentistry

Department: Paediatric and Pre	Specification	Quanti	ty	Quantity	Adequate / Inadequate
ntal Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, and motorised suction, micromotor attachment with contra angle miniature hand piece, airotor attachment with miniature hand piece, dental operater stool (40% dental chairs shall be pedo chairs)	graduate student Two chairs unit for faculty	st- and	23 (9 pedo+1 Left handed)	Adequate
		1 Unit 2	Units		
		3	4	4	Adequate
edo extraction forceps sets utoclaves for bulk instrument sterilization		1	2	2	Adequate
acuum (Front loading)		1	1	1	Adequate
VG with intra oral x-ray unit		1	1	1	Adequate
automatic developer		2	3	4	Adequate
Pulp tester		1	1	3	Adequate
Apex locator		1	1	18	Adequate
Rubber dam kit	One set per student	-	1	2	Adequate
njectable GP condenser		1	1	2	Adequate
Endodontic pressure syringe		1	4	8	Adequate
Glass bead steriliser		2	3	3	Adequate
Spot welder		2	4	15	Adequate
Ultrasonic scalers		2	1	3	Adequate
Needle destroyer		1 1	1	6	Adequate
Formalin chamber		1	1	3	Adequate
Ultrasonic cleaner capacity 3.5 lts		2	3	9	Adequate
X-ray viewer		1	2	4	Adequate
Amalgamator		2	2	2	Adequate
Plaster dispenser		1	2	2	Adequate
Dental lathe		2	3	3	Adequate
Vibrator Typodonts	One set per student	1	1	18 pedo + Permane	18 Adequate
**		1	1	1	Adequate
Soldering unit		2 Sets	2 Se	ets 3 sets	Adequate
Band pinching beak pliers		2	3	6	Adequate
Proximal contouring pliers		2	3	3	Adequate
Crown crimping pliers Double beak pliers anterior and posterior		2	3	3	Adequate
		2	3		Adequate
Lab micro motor		1	2		Adequate
Acryliser		1	1	9	Adequate
Magnifying loupes Conscious sedation unit	Desirable	1	1	1	Adequate
		1	1		Adequate
Pulse oxymeter Phantom head table with attached Light, Airotor and micro motor	One set per each P.G. Student	1	1		Adequate
Computer with internet connection with attached printer and scanner	Madela	1	1	2	Adequate
	Made	1	1	1	Adequate
LCD projector Refrigerator	DEAN	1		1 1	Adequate

Department: Oral Medicine and Radiology

Name	Specification	unit per post-		Available Quantity	Adequate /
Dental Chairs and Units	Electrically operated with shadowless lamp, spittoon, 3 way syringe, instrument tray and suction			15	Adequate
		1 Unit	2 Units		
RVG with intra oral	55-70 Kvp with digital	1	1	1	Adequate
radiography machine (FDA Approved)	compatibility	, "i			
Extra oral radiography machine	100 kvp	1	1	1	Adequate
Panoramic radiography (OPG) machine with cephalometric and TMJ attachment with printer	Digital compatibility	1	1	1	Adequate
Intra-oral camera		1	2	1	Adequate
Pulp tester		2	4	2	Adequate
Autoclave		1	1	2	Adequate
Punch biopsy tool		2	3	2	Adequate
Biopsy equipment		1	2	1	Adequate
Surgical trolley		2	2	2	Adequate
Emergency medicines kit		1	1	1	Adequate
Extra oral cassettes with intensifying screens (Conventional and rare earth)		4	6	4	Adequate
Lead screens		2	2	2	Adequate
Lead aprons		2	2	2	Adequate
Lead gloves		2	2	2	Adequate
Radiographic filters (Conventional and rare earth)	BELLEVIEW .	1	1	1	Adequate
Dark room with safe light facility		1	- 1	1	Adequate
Automatic radiographic film processors		2	2	2	Adequate
Radiographic film storage lead containers		1	1	1	Adequate
Thyroid collars		1	1	1	Adequate
Digital sphygmomanometer		1.	1	1	Adequate
Digital blood glucose tester		1	1	1	Adequate
Digital camera		1	1	1	Adequate
X-ray viewer boxes		2	3	2	Adequate
Lacrimal probes		2 sets	2 Sets	2	Adequate
Sialography cannula		2 sets	2 Sets	2	Adequate
Illuminated mouth mirror and probe		2	2	2	Adequate
Computer with internet connection with attached printer and scanner		1	1	1	Adequate
LCD projector		1	1	1	Adequate
Refrigerator		1	1	1	Adequate

Dean / Principal Stamp & Signature

Department: Oral & Maxillofacial Pathology and Oral Microbiology

Name	Specification	Quantity		Available Quantity	Adequate / Inadequate
		1 Unit 2 Units			
Dental Chairs and Units	Electrically operated with shadow less lamp, spittoon, 3 way syringe, instrument tray and suction	3	6		Adequate
Adequate laboratory glassware's as required for processing of biopsy specimens and staining.	Reasonable quantit y should be made available				Adequate
Adequate tissue capsules tissue embedding cassettes	Reasonable quantity should be made available			Available	Adequate
Paraffin wax bath	thermostatically controlled	1	1	1	Adequate
Leuckhart pieces		10	10	10	Adequate
Block holders		25	25	25	Adequate
Microtome	Manual	1	1	1	Adequate
Microtome	semi – automated	1	1	- 1	Adequate
Tissue floatation water bath	thermostatically controlled	1	1	1	Adequate
Slide warming table		1	1	1	Adequate
Steel slide racks for staining		5	5	7	Adequate
Diamond glass marker		2	2	2	Adequate
Research microscope with phase contrast, dark field, polarization, image analyzer, photomicrography attachments		1	1	1	Adequate
Multi head microscope	Penta headed	1	1	2	Adequate
Binocular compound microscope		2 for faculty and one per student	4 for faculty and one per student		Adequate
Stereo microscope		1	1	1	Adequate
Aluminum slide trays		5	5	5	Adequate
Wooden / plastic slide boxes		5	5	5	Adequate
Wax block storing cabinet		у	10,000 capacit y	capacity available	e
Slide storing cabinet		5,000 capacit	10,000 capaci		
Refrigerator		1	1	1	Adequate
Pipettes		5	5	-5	Adequate
Surgical kit for biopsy		3	6	3	Adequate
Immuno histo chemistry lab	1000	1	1	1	Adequate
Computer with Internet Connection with attached printer and scanner	DEAN Y.M.T. Dental	1	1	2	Adequate

& Hospital Kharghar, Navi Mumba' - 410 210

Desirable Equipment				
Cryostat	1	1		-
Fluorescent microscope		1	1	Adequate
Hard tissue microtome	1	1	-	
Tissue storing cabinet (frozen)	1	1		
Microwave	1	1	1	Adequate

MIVandeka

Department: Orthodontics and Dentofacial Orthopedics

ame	Specification Quant		Specification Quant		Quantity		Adequate / Inadequat e
Dental Chairs and Units	Electrically operated with shadow less lamp, spittoon, 3 way syringe, instrument tray and motorized suction	One chair & unit per PG student and Two chairs with unit for the faculty		20	Adequate		
		1 Unit	2 Units				
/acuum /pressure moulding unit		1	1	1	Adequate		
Hydrogen soldering unit		1	1	1	Adequate		
ab micromotor		3	5	13	Adequate		
Spot welders		3	5	20	Adequate		
Model trimmer (Double disc)		2	3	3	Adequate		
Light curing unit		2	2	2	Adequate		
High intensity light curing unit		1	2	15	Adequate		
Polishing lathes		2	3	3	Adequate		
Fracing tables		3	5	11	Adequate		
SLR digital camera		1	1	1	Adequate		
Scanner with transparency adapter		1	1	1	Adequate		
X-ray viewer		3	4	4	Adequate		
LCD projector		1	1	1	Adequate		
Autoclaves for bulk instrument Sterilization vacuum (Front loading)		1	1	5	Adequate		
Needle destroyer		1	1	2	Adequate		
Dry heat sterilizer		1	1	2	Adequate		
Ultrasonic scaler		1	1	14	Adequate		
Sets of Orthodontic pliers		3	3	11	Adequate		
Orthodontic impression trays		3	5	150 SINGLE			
Ultrasonic cleaner capacity 3.5 Its		1	1	1	Adequate		
Electropolisher		1	1	1	Adequate		
Typodonts with full teeth set		3	3	16	Adequate		
Anatomical articulator with face box	W	1	1	3	Adequate		
Free plane articulators		1	1	4	Adequate		
Hinge articulators		4	4	12	Adequate		
Computer software for cephalometrics		1	1	1	Adequate		
Computer with internet connection with attached printer and scanner		1	1	1	Adequate		
Refrigerator		1	1	1	Adequate		

D:\LIC_2023-24\LIC_Forms\LIC_Form_Dental_25012023.docx